

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.13 revision 0  
Project number 3788297  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Manufacturer

Mettler-Toledo GmbH  
Im Langacher 44  
8606 Greifensee  
Switzerland

Identification of the certified type

**A Non-automatic weighing instrument**  
Type : FreshStar...

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
28 June 2024

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.13 revision 0  
Project number 3788297  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3788297-01 dated 28 June 2024 that includes 25 pages;
- No. NMI-3788297-02 dated 28 June 2024 that includes 21 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing ranges	Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)	$n \leq 3000$	
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -\text{Max}$	
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	100 – 240 V AC 50/60 Hz	
Application	Intended to be used for direct sales to the public	
Software identification	Scale application	SA1.y.z(y: 0-99, z: 0-9999)
	Viewer application	VA1.y.z(y: 0-99, z: 0-9999)

The procedure to display/print the software identification and to recall the event logger is as follows:

- Press 'Menu' -> Enter password 246813-> 'Configuration' -> 'System' -> 'Printout System Info'.

This will result in a printout with all the information related to the instrument and software; Software identification for scale application and viewer application are also shown at start-up.

### Revision History

Revision	Date	Change(s)
0	2024-06-28	Initial issue.