O I M L   C e r t i f i c a t e

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D. Schmidt
Manufacturer	Mettler-Toledo GmbH Im Langacher 44 8606 Greifensee Switzerland
Identification of the certified type	A **Non-automatic weighing instrument** Type : FreshStar…
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class ⟨III⟩

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
17 February 2025

Certification Board

OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.13 revision 1
Project number 3835046
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated reports:
- No. NMi-3788297-01 dated 28 June 2024 that includes 25 pages;
- No. NMi-3788297-02 dated 28 June 2024 that includes 21 pages.
- No. NMi-3835046-01 dated 17 February 2025 that includes 28 pages;
- No. NMi-3835046-02 dated 17 February 2025 that includes 14 pages;
- No. NMi-3835046-03 dated 17 February 2025 that includes 21 pages;
- No. NMi-3835046-04 dated 17 February 2025 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		(III)
Maximum capacity		$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval		$e \geq 1 \text{ g}$
Weighing range(s)		Single interval Multi-interval
Maximum number of scale intervals (one weighing range)		$n \leq 3000$
Maximum number of scale intervals (multi-interval)		$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges		2
Tare		$T \leq -\text{Max}$
Temperature range		-10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for direct sales to the public
Software identification	Scale application (Windows)	WSAx.y.z (x=0..9, y=0..99, z=0..9999)
	Viewer application (Windows)	WVAa.b.c (a=0..9, b=0..99, c=0..9999)
	Scale application (MTSmartOS)	SA1.y.z(y=0..99, z=0..9999)
	Viewer application (MTSmartOS)	VA1.y.z(y=0..99, z=0..9999)

The procedure to display/print the software identification and to recall the event logger is as follows:
- Press 'Menu' -> Enter password 246813-> 'Configuration' -> 'System' -> 'Printout System Info'. This will result in a printout with all the information related to the instrument and software.
- Software identification for scale application and viewer application are also shown at start-up.

OIML Certificate

Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
0	2024-06-28	Initial issue.
1	2025-02-17	Addition of hanging type scale and alternative main board with updated software for Windows.