

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.17 revision 1  
Project number 3681702  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

CAS Corporation  
# 19 Ganap-Ri, Gwangjuk-Myun  
Yangju City, Kyunggi-Do  
South Korea

Identification of the  
certified type

**A Non-automatic weighing instrument**  
Type : EBW-xxxAD, EBW-xxxAS

x can be a number or letter and blank, representing the maximum capacity.

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76-1:2006** for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
3 October 2024

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-24.17 revision 1  
Project number 3681702  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3681702-01 revision 1 dated 3 October 2024 that includes 50 pages;
- No. NMI-3681702-02 revision 1 dated 3 October 2024 that includes 33 pages;
- No. NMI-3681702-03 dated 23 August 2024 that includes 19 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	15 kg ≤ Max ≤ 150 kg
Verification scale interval	e ≥ 2 g
Weighing range	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	n ≤ 7500
Maximum number of scale intervals (multi-interval)	n ≤ 3000 (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	T ≤ -100% for instruments with one weighing range
Temperature range	-10 °C / + 40 °C
Power supply voltage	100-240 V AC 50/60 Hz to 9 V DC adapter 7,3 V DC rechargeable battery
Application	Intended to be used for direct sales to the public
Software identification Version number	V 5xx or yy5xx with non legally relevant parts: xx = any character reflecting software changes, and yy = any character reflecting the country code (y, x = 0... 99)

The software identification is displayed at start-up.

### Revision History

Revision	Date	Change(s)
0	2024-08-23	Initial issue.
1	2024-10-03	Correction of reports NMI-3681702-01 and NMI-3681702-02, renamed to NMI-3681702-01 revision 1 and NMI-3681702-02 revision 1.