

OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.22 revision 0
Project number 3750240
Page 1 of 2

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Manufacturer

ArjoHuntleigh AB
Hans Michelsengatan 10
211 20 Malmö
Sweden

Identification of the certified type

A Non-automatic weighing instrument
Type : Maxi Move 5

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
17 October 2024

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-24.22 revision 0
Project number 3750240
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMI-3750240-01 dated 17 October 2024 that includes 44 pages;
- No. NMI-3750240-02 dated 17 October 2024 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$227 \text{ kg} \leq \text{Max} \leq 272 \text{ kg}$
Verification scale interval	$e \geq 200 \text{ g}$
Weighing ranges	Single interval
Maximum number of scale intervals	$n \leq 1360$
Tare	$T \leq -\text{Max}$
Temperature range	$+10 \text{ °C} / + 40 \text{ °C}$
Power supply voltage	24 V DC by the battery pack of the lifter
Application	To be used only for weighing patients
Software identification	2.0.z (z = 0...255 and represents changes to the non-legally relevant part of the software)

The scale software identification is displayed after holding the left and right selection buttons

simultaneously for two seconds and then select **i**.

Revision History

Revision	Date	Changes
0	2024-10-17	Initial issue.