



# OIML Certificate

#### **OIML Member State** The Netherlands



Number R76/2006-A-NL1-24.29 revision 0 Project number 3806248 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Applicant and Ishida Co., Ltd

Manufacturer 44 Sannocho, Shogoin, Sakyo-ku

Kyoto, 606-8392

Japan

Identification of the

certified type

A Non-automatic weighing instrument

IPC / s-boxR

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76-1:2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



#### Issuing Authority

### NMi Certin B.V., OIML Issuing Authority NL1 6 January 2025



**Certification Board** 

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







NMi Certin B.V. Thiissewea 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl







# OIML Certificate

# **OIML Member State** The Netherlands



Number R76/2006-A-NL1-24.29 revision 0 Project number 3806248 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated reports:

- No. NMi-3806248-01 dated 6 January 2025 that includes 47 pages;
- No. NMi-3806248-02 dated 6 January 2025 that includes 19 pages;
- No. NMi-3806248-03 dated 6 January 2025 that includes 18 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	or (III)
Maximum capacity	1,5 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 1 g
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	n ≤ 1500
Maximum number of scale intervals (multi-interval)	n ≤ 1500 (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \le -50\%$ for instruments with one weighing range $T \le -Max_1 - e_1$ for multi-interval instruments
Temperature range	-5 °C / +40 °C
Power supply voltage	230 V AC 50/60 Hz 3 V DC by battery
Software Version number	L004

The software identification is displayed at start-up.



The non-automatic weighing instrument has embedded software.

### **Revision History**

Revision	Date	Change(s)
0	2025-01-06	Initial issue.



 $\mathbf{T}$